

The following table outlines the energy protocol recommendations for the Philips MRx Biphasic Monitor/Defibrillator/Pacer

Arrhythmia	Shock sequence
VF/non-perfusing VT	150, 150, 150 J
Perfusing VT/PSVT (synchronized)	100, 150, 200 J
Atrial Fibrillation (synchronized)	100, 150, 200, 200 J or 150, 200, 200 J
Atrial Flutter (synchronized)	50, 100, 150 J
Intrathoracic defibrillation	10, 20, 30 J
Pediatric defibrillation	1-2 J/Kg, 2-3 J/Kg, 3-4 J/Kg
Pediatric synchronized cardioversion	0.5 - 1 J/Kg, 1-2 J/Kg